

i-STAT 1 SYSTEM MANUAL CONFIGURATION

Please ensure that the contents of your System Manual are complete and up to date. In the event that your System Manual does not contain the current configuration, it is recommended that you contact your i-STAT support provider.

As of April 2025, your i-STAT 1 System Manual should be configured with the contents as listed below and in the order shown.

ITEM	Art #
Cover Sheet	714336-00S
Table of Contents.....	714362-00AK
Section 1.....	714363-00AG
Section 2.....	714382-00H
Section 3.....	714364-00AA
Section 4.....	714365-00L
Cartridge and Test Information.....	714258-00W
Section 5.....	714367-00J
Section 6.....	714368-00Q
Section 7.....	714369-00O
Section 8.....	714372-00T
Section 9.....	714373-00M
Section 10.....	714374-00R
Section 11.....	714381-00O
Section 12.....	714376-00AC
Section 13.....	714377-00X
Section 14.....	714378-00K
Section 15.....	714371-00M
Section 16.....	714379-00L
Section 17.....	714380-00S

Technical Bulletins

Sample Type Customization on the i-STAT 1 Analyzer.....	765893-00C
Instructions for Restoring Analyzers that Produce *** for Hematocrit and Quality Check Code 23.....	721215-00D
K ₂ EDTA and K ₃ EDTA Customization for Hematocrit on the i-STAT System.....	716240-00F
ACT Test Result Options: Prewarmed vs. Non-Prewarmed Result Calibration Modes for the i-STAT 1 Analyzer.....	715617-00F

i-STAT Celite ACT and i-STAT Kaolin ACT	
Heparin Linearity Procedure.....	714547-00H
Support Services.....	716144-00AM
The i-STAT System and Waived Status.....	731678-00F
The Presence of Latex Rubber in i-STAT	
System Components.....	721296-00I